



Sulfur Dioxide
Breakthrough Indicator Sticker
(SO₂ BTIS)
(PN: 168-0000)



Manual



Sulfur Dioxide Breakthrough Indicator Sticker Manual

1. Application

The Sulfur dioxide BTIS (PN: 167-0000) is qualitative (yes/no) colorimetric indicator for the saturation and end-of-service life of ductless hood filters. The indicator is designed to provide real-time indication of the breakthrough of Sulfur dioxide.

2. Specifications

- | | |
|------------------------------|--|
| a. Weight: | 0.4g (0.02oz) |
| b. Dimensions: | 2.8mm (0.11in), ϕ : 31.8mm (1.25in) |
| c. Operating temperature: | -20°C to 55°C (-4°F to 131°F) |
| d. Operating humidity: | 5% RH to 85%RH |
| e. Minimum detectable limit: | 0.5ppm-hr at 30 cm/sec face velocity |
| f. Color change: | Yellow to orange to dark red |
| g. Storage temperature: | 4°C to 25°C, (39°F to 77°F) |
| h. Shelf life: | 1 year at 4°C to 25°C, (39°F to 77°F) |
| i. Service life: | 14 months |

Cross interferences: Basic vapors in high concentrations impair the performance of the sulfur dioxide breakthrough indicator. Acid vapors produce similar color. No other interferences are known.

3. Operating Instructions

- Ensure that packaging pouch is intact.
- Open packaging pouch by tearing off the top part from one of side notches
- Remove indicator sticker from the packaging pouch.
- Peel off the protective liner to expose the bottom adhesive (Figure 1).



Caution: Do not touch bottom adhesive or the exposure area

- Hold the sticker from the edges, as shown in Figure 2, and place it on center clean area of the filter's outlet with the reading area (glossy surface) of the sticker facing up.
- Press firmly to attach sticker to the filter's outlet (Figure 3).
- Replace filter when the reading area of the indicator changes color to orange or dark red.




Figure 1



Figure 2



Figure 3

CHEMTEQ[®]

Chemteq, Inc., 600 West 24th Street, Suite B, Norfolk, Virginia 23517, Tel 757-622-2223, sales@chemteq.net

©2009 Chemteq, Inc.

www.chemteq.net

MP-1688100-091220-B