


## Basic Vapors Breakthrough Indicator Sticker Type B (BTIS B) (PN: 254)

*Manual*



## 1. Application

The Basic Vapors breakthrough indicator sticker type B (PN: 254) is a colorimetric indicator for the detection of traces of basic vapors in air, suitable locations are the inside of window glass of chemical transport vehicles, and the inside window of chemical storage cabinets where representative samples can be obtained.

## 2. Specifications

- |                              |   |
|------------------------------|---|
| a. Weight:                   | 0.4g (0.02oz)                           |
| b. Dimensions:               | 2.0mm (0.11in), $\phi$ : 38.1mm (1.5in) |
| c. Operating temperature:    | -20°C to 65°C (-4°F to 149°F)           |
| d. Operating humidity:       | 5% RH to 85%RH                          |
| e. Minimum detectable limit: | 4ppm-hr at 30 cm/sec face velocity      |
| f. Color change:             | Orange to green to blue                 |
| g. Storage temperature:      | 4°C to 25°C, (39°F to 77°F)             |
| h. Shelf life:               | 1 year at 4°C to 25°C, (39°F to 77°F)   |
| i. Service life:             | 4 months                                |

Cross interferences: Acidic vapors in high concentrations impair the performance of the Bases breakthrough indicator. No other interferences are known.

## 3. Operating Instructions

- Ensure that packaging pouch is intact.
- Open packaging pouch by tearing off the top part from one of side notches.
- Remove indicator sticker from the packaging pouch (Figure 1).
- Peel off the protective liner to expose adhesive (Figure 2).



**Caution: Avoid touching adhesive and exposure area**

- Figure 3 shows the view result area of the indicator and the locations of the three adhesive discs.
- Hold the indicator from the edges and place it on a suitable location position on the inside of window glass of chemical transport vehicles or the inside window of chemical storage cabinets where representative samples can be obtained.
- Press firmly to adhere the indicator on the location surface.
- Formation of green or blue color indicates the presence of basic vapors.



Figure 1



Figure 2




Figure 3

**CHEMTEQ**<sup>®</sup>