

CHEMTEQ[®]

Organic Vapors
Breakthrough Indicator
With Auxiliary Filter Trap
OV BTI AFT

(PN: 666)



Manual

CHEMTEQ[®]

600 West 24th Street, Suite B
Norfolk, Virginia 23517, USA

Tel: 757-622-2223

Toll-free: 855-CHEMTEQ (855-243-6837)

sales@chemteq.net



1. Application

The OV BTI AFT (PN: 666) is qualitative (yes/no) colorimetric indicator for the saturation and end-of-service life of filters. The BTI designed to provide real-time indication of the breakthrough of organic vapors. The indicator is most suited for process vents filters associated with distillation, thin-film evaporation, solvent extraction, waste containers and liquids transfers. The indicator is equipped with auxiliary filter trap to ensure no organic vapors escape to the outside environment while the indicator changing color. The indicator is designed to provide real-time indication of the breakthrough of organic vapors, including: Acetone, Acetonitrile, Acids, Acrylonitrile, Aliphatic hydrocarbons (i.e., hexane), Aromatic Hydrocarbons (i.e., benzene, toluene and xylenes), Chlorinated hydrocarbons (i.e. carbon tetrachloride, chloroform and dichloromethane (methylene chloride), Ethanol, Ethyl Acetate, Ethyl acrylate, Ethyl ether, Gasoline, HFIP (Hexafluoroisopropanol), Methanol, Methyl acrylate, Naphtha, Phenol, Sulfolane, THF (Tetrahydrofuran). This indicator cannot detect basic organic compounds such as pyridines, aliphatic and aromatic amines. Please contact us for alternative indicator for detecting basic organic compounds.

2. Specifications

2.1. Overall Specifications

- a. Weight: 118g (4.2oz)
- b. Dimensions:
 - Breakthrough indicator (Part A): 89.9mm (3.5in), diameter 24.5mm (1.0in)
 - Auxiliary filter trap (Part B): 41mm (1.6in), diameter 81mm (3.2in)
- c. Inlet dimensions: ½" MNPT
- d. Operating temperature: 4°C to 32°C (39°F to 89.6°F)
- e. Operating humidity: 5% RH to 85%RH
- f. Minimum detectable limit: See Performance Specification 2.
- g. Color change: Orange to red
- h. Storage temperature: 4°C to 25°C, (39°F to 77°F)
- i. Shelf life: 12 months at 4°C to 25°C, (39°F to 77°F)
- j. Service life: 12 months

2.2. Performance Specification

To determine the sensitivity of the breakthrough indicator, a solution/mixture of 10% solvent in water was bubbled with ambient air at a flow rate of 5cc/min. The airflow passed through the breakthrough indicator until a color change was observed. The elapsed time to observe the first noticeable and the final colors for the respective organic solvent is depicted in the table below.

3. Instructions

- a. Ensure that packaging pouch is intact.
- b. Open packaging pouch by tearing off the top part from one of side notches.
- c. Remove the breakthrough indicator (Part A), Figure 1, and the auxiliary filter trap (Part B), Figure 2, from the packaging pouch.
- d. Screw the breakthrough indicator (Part A) into the auxiliary filter trap (part B) as shown in Figure 3.
- e. Remove the protective red plug to activate the breakthrough indicator.
- f. Screw in the Breakthrough Indicator into the ½" threaded carbon absorber outlet lid. Teflon tap can be used on threads to ensure proper seal.



Caution: Only hand tighten indicator into carbon absorber

- g. Replace carbon absorber when the Breakthrough Indicator changes color to red.

Solvent (10% in Water)	Breakthrough Indication Time	
	First Noticeable Color (min)	Final Color (min)
Acetone	10	30
Acetonitrile	10	30
Benzene	4	43
Carbon tetrachloride	10	30
Chloroform	2	8
Dichloromethane (methylene chloride))	2	8
Ethanol	10	30
Ethyl Acetate	2	5
Gasoline	14	60
Hexane	10	30
HFIP Hexafluoroisopropanol	1	4
Methanol	2	5
Naphtha	10	30
Phenol	45	6 hours
Sulfolane	8	17
THF (Tetrahydrofuran)	10	20
Toluene	2	8
Trifluoroacetic acid	4	30
Xylenes	14	60

COLORS ARE APPROXIMATE



Figure 1



Figure 2



Figure 3